



Industrial Temperature Sensors Ltd.

Unit 18, Nass Industrial Estate, Naas, Co. Kildare

Tel: 045-898164 Fax: 045-896521 Web: www.itsirl.com email: info@itsirl.com

CERTIFICATE OF CALIBRATION

Certificate No: C208323/C1

Customer: Rilta Enviromental

Order No.:3530

Instrument: Resistance Thermometer Insert & Tx's

Model:PR5331B3B1

Serial:970249650

Range: 0+100C

Test Equipment:

(1) AMETEK DTI 1000 S/N. 013.02-01-950702 CERT No. 58747
AMETEK PT100 S/N. 515361-01 CERT No. 58747

(2) AMETEK Macal mA LOOP CALIBRATOR
S/N. 008653-00239 CERT No. 58751

REFERENCE TEMPERATURE	INSTRUMENT READING	
	MA	ERROR MA
0.001C	3.995	-0.005
50.002C	12.000	0
99.960	19.990	-0.01

TESTED BY: STEPHEN RALPH

Date: 19/02/2008



Industrial Temperature Sensors Ltd.

Unit 18, Nass Industrial Estate, Naas, Co. Kildare

Tel: 045-898164 Fax: 045-896521 Web: www.itsirl.com email: info@itsirl.com

CERTIFICATE OF CALIBRATION

Certificate No: C208323/C2

Customer: Rilta Enviromental

Order No.:3530

Instrument: Resistance Thermometer Insert & Tx's

Model: PR5331B3B1

Serial: 970249643

Range: 0+100C

Test Equipment:

(1) AMETEK DTI 1000 S/N. 013.02-01-950702 CERT No. 58747
AMETEK PT100 S/N. 515361-01 CERT No. 58747

(2) AMETEK Macal mA LOOP CALIBRATOR
S/N. 008653-00239 CERT No. 58751

REFERENCE TEMPERATURE	INSTRUMENT READING	
	MA	ERROR MA
DEGREE C		
0.001C	3.996	-0.004
50.002C	11.990	-0.01
99.960	19.990	-0.01

TESTED BY: STEPHEN RALPH

Date: 19/02/2008