

Accreditation Certificate

Cork County Council

Wastewater Testing Laboratory, Inniscarra, Co. Cork

Testing Laboratory

Registration number: 016T

is accredited by the Irish National Accreditation Board (INAB) to undertake testing as detailed in the Schedule bearing the Registration Number detailed above, in compliance with the International Standard ISO/IEC 17025:2005 2nd Edition "General Requirements for the Competence of Testing and Calibration Laboratories" (This Certificate must be read in conjunction with the Annexed Schedule of Accreditation)

Date of award of accreditation: 01:10:2002

Date of last renewal of accreditation: 20:09:2007

Expiry date of this certificate of accreditation: 01:10:2012

This Accreditation shall remain in force until further notice subject to continuing compliance with INAB accreditation criteria, ISO/IEC 17025 and any further requirements specified by the Irish National Accreditation Board.

Manager: Jom Dempoy

Dr Máire Walsh


Mr Tom Dempsey

Issued on 23 June 2008

Organisations are subject to annual surveillance and are re-assessed every five years. The renewal date on this Certificate confirms the latest date of renewal of accreditation. To confirm the validity of this Certificate, please contact the Irish National Accreditation Board.

The INAB is a signatory of the European co-operation for Accreditation (EA) Testing Multilateral Agreement (MLA) and the International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Arrangement.

Wilton Park House, Wilton Place, Dublin 2, Ireland. Tel +353 1 607 3003 Fax +353 1 607 3109 E-mail inab@inab.ie Web www.inab.ie



Wilton Park House, Wilton Place, Dublin 2, Ireland Tel +353 1 607 3003 Fax +353 1 607 3109

Schedule of Accreditation



(Annex to Accreditation Certificate)

Permanent Laboratory: Category A

CORK COUNTY COUNCIL

Chemistry Testing Laboratory

Initial Registration Date :

25-April-1991

Postal Address:

Waste Water Laboratory

(Address of other locations

Inniscarra

as they apply)

Co. Cork

Telephone:

+353 (21) 4532700

Fax:

+353 (21) 4532777


E-mail:

Contact Name:

Ms M Cherry

Facilities:

Normally not available for Public testing



Wilton Park House, Wilton Place, Dublin 2, Ireland Te(+353 1 607 3003 Fax +353 1 607 3109

Schedule of Accreditation



Permanent Laboratory: Category A

THE IRISH NATIONAL ACCREDITATION BOARD (INAB) is the Irish body for the accreditation of organisations including laboratories.

Laboratory accreditation is available to testing and calibration facilities operated by manufacturing organisations, government departments, educational institutions and commercial testing/calibration services. Indeed, any organisation involved in testing, measurement or calibration in any area of technology can seek accreditation for the work it is undertaking.

Each accredited laboratory has been assessed by skilled specialist assessors and found to meet criteria which are in compliance with ISO/IEC 17025 or ISO/IEC 15189 (medical laboratories). Frequent audits, together with periodic inter-laboratory test programmes, ensure that these standards of operation are maintained.

Testing and Calibration Categories:

Category A:

Permanent laboratory calibration and testing where the laboratory is erected on a fixed

location for a period expected to be greater than three years.

Category B:

Site calibration and testing that is performed by staff sent out on site by a permanent

laboratory that is accredited by the Irish National Accreditation Board.

Category C:

Site calibration and testing that is performed in a site/mobile laboratory or by staff sent out by such a aboratory, the operation of which is the responsibility of a permanent

laboratory accredited by the Irish National Accreditation Board.

Category D:

Site calibration and testing that is performed on site by individuals and organisations that do not have a permanent calibration/testing laboratory. Testing may be performed using

(a) portable test equipment

(b) a site laboratory

(c) a mobile laboratory or

equipment from a mobile or site laboratory

Standard Specification or Test Procedure Used:

The standard specification or test procedure that is accredited is the issue that is current on the date of the most recent visit, unless otherwise stated.

Glossary of Terms

Facilities:

Public calibration/testing service:

Commercial operations which actively seek work from others.

Conditionally available for public calibration/testing:

Established for another primary purpose but, more commonly than not,

is available for outside work.

Normally not available for public

Unavailable for public calibration/testing more often than not.

calibration/testing:

Laboratory users wishing to obtain assurance that calibration or test results are reliable and carried out to the Irish National Accreditation Board criteria should insist on receiving an accredited calibration certificate or test report. Users should contact the laboratory directly to ensure that this scope of accreditation is current. INAB will, on request, verify the status and scope.



Cork County Council Chemical Testing Laboratory

Permanent Laboratory:

Category A

(P9)	lassification number	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
766	Waters	Chemical analysis:	Documented in-house methods based on
			Standard Methods for the Examination of Water
.01	Waters for		& Wastewater 21 st Edition APHA (See Note 1)
	domestic purposes	Biochemical Oxygen Demand	No. 1 Membrane electrode
	Surface and ground	2 - 145,000 mg/l	
	waters	See of for	
		pH Ruffective	CP No. 5 Electrometry
		2 - 12 eciton de la companya del companya del companya de la compa	
		pH 2 - 12 Suspended Solids 0.5 - 17,500 mg/l	CP No. 3 Gravimetric
		Chemical Oxygen Demand	CP No. 6 Reflux - colourmetric method
		21 - 135 mg/l	CONTROL OF NOTICE
		120 - 670,000 mg/l	
		, 120 eve, 100 mg, 1	·
		Total phosphorus	US-EPA Approved method/HACH
		0.2 - 5,300 mg/l	Method CP No.20
		Ammonia	Documented in-house method CP22 by Konelab
		0.1 - 1,000 mg/l NH ₃ - N	based on Method for the Examination of Waters
<u> </u> 			and
			Associated Material HMSO:1981
	· · · · · · · · · · · · · · · · · · ·		



Cork County Council Chemical Testing Laboratory

Permanent Laboratory: Category A

	(P9)	assification number ls/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
	766	Waters		
	.01	Waters for	Orthophosphate as P (Konelab)	CP No. 23 Ascorbic Acid Method
	.01	domestic purposes	Range: 0.005-1.00 mg O-PO4 P/L	CF NO. 23 ASCOIDIC ACIG MECHOG
			۶,	&°.
		Surface and ground	High Range: 1000 mg 0-P04 P/L	
ļ		waters	Method Detection Limit: 0.02 mg 0.PQ.P/L	
			Chloride (Konelab)	CP No. 24 Ferricyanide Method
			Range: 25-250 mg/L-Cl	
			High Range Conc.: 86,000 mg/L Cl-	
			Method Detection Limit: 25 mg/L Cl-	
-		-	ation	
			Sulphate (Konelab)	CP No. 25 Documented in-house method by
			Range: 30-250 mg/L SO4/L	Konelab based on method for the examination
			High Range Conc.: 35,000 mg/L SO4/L	of waters and waste waters and associated
			Method Detection Limit: 30 mg SO4/L	material HMSO: 1981
Ì			· · ·	
.				
İ				
			•	
				:



Cork County Council

Permanent Laboratory:

Category A

Chemical Testing Laboratory

P9)	lassification number als/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
766	Waters	Chemical analysis	Documented in-house methods based on Standard
			Methods for the Examination of Water&
.05	Trade Wastes		Wastewater 21 st Edition APHA (See Note 1)
	Industrial effluents	Biochemical Oxygen Demand	CRNo. 1 Membrane electrode
	Urban Wastewater	2 - 145,000 mg/l	i and
	Municipal Wastewater	obside the state of the state o	
	•	Biochemical Oxygen Demand 2 - 145,000 mg/l pH 2 - 12 tot its petion bill owner bedricht owne	CP No. 5 Electrometry
		2-12 gedioniei	
		COL ITELLE	
		tooly.	
		Suspended Solids	CP No. 3 Gravimetric
		0.5 - 19,500 mg/l	
		Chemical Oxygen Demand	CP No. 6 Reflux - colourmetric method
		21 - 135 mg/l	
		120 - 670,000 mg/l	·
		, ,	
		Total phosphorus	US-EPA Approved method/HACH
		0.2 - 5,300 mg/l	Method CP No.20
		Ammonia	Documented in-house method CP22 by Konelab
		0.1 - 1,000 mg/l NH3-N	based on Method for the Examination of Waters
			and Associated Material HMSO: 1981.



Cork County Council

Permanent Laboratory:

Category A

Chemical Testing Laboratory

P9)	lassification number als/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
'66	Waters	Chemical analysis	Documented in-house methods based on Standar
			Methods for the Examination of Water&
05	Trade Wastes		Wastewater 21 st Edition APHA (See Note 1)
	Industrial effluents	3	CP No. 1 Membrane electrode
	Urban Wastewater	obsessor and	
	Municipal Wastewater	att att	
		Orthophosphate as P. (Konetab)	CP No. 23 Ascorbic Acid Method
		Range: 0.005 - 1000 mg 0-PO4 P/L	
		High Ranges 1000 mg O-PO4 P/L	
		Method Detection Limit: 0.02 mg O-	
		PO4 PAE	
		Coll	·
		Chloride (Konelab)	CP No. 24 Ferricyanide Method
		Range: 25-250 mg/L Cl-	
		High Range Conc.: 86,600 mg /L Cl-	
		Method Detection Limit: 25mg / L Cl-	
		Sulphate (Konelab))	CP No. 25 Documented in-house method by
		Range: 30-250 mg/L SO4 /L	Konelab based on method for the examination o
		High Range Conc.: 35,000 mg/L SO4 /L	waters and waste waters and associated materia
		Method Detection Limit: 30 mg SO4 /L	HMSO: 1981
			·

Consent of copyright owner reduced by any other use.